

This declaration is issued under the sole responsibility of the manufacturer INFICON. The object of the declaration is to certify that this equipment, designed and manufactured by:

## INFICON Inc. Two Technology Place East Syracuse, NY 13057 USA

is in conformity with the relevant Community harmonization legislation. It has been constructed in accordance with good engineering practice in safety matters in force in the Community and does not endanger the safety of persons, domestic animals or property when properly installed and maintained and used in applications for which it was made.

Equipment Description:	GAS-Mate <sup>®</sup> Combustible Gas Leak Detector	
Model Number:	718-202-Gxx (App	licable to all Group numbers)
Applicable Directives:	2014/30/EU 2011/65/EU	General EMC as amended by 2015/863/EU RoHS
Applicable Standards: Safety:	IEC 61010-1:2010	Safety requirements for electrical equipment for measurement, control, and laboratory use. General requirements
Emissions:	EN 61326-1:2013	Edition 2.0 (Radiated, Conducted & Harmonic Emissions) (EMC- Measurement, Control & Laboratory Equipment)
Immunity:	EN 61326-1:2013	Edition 2.0 (EMC – Measurement, Control & Laboratory Equipment) Immunity per Table A.1 – Portable Test and Measurement Equipment
RoHS	Compliant	
CE Implementation Date:	August 31, 2021	

Bran N.

Authorized Representative: Brian King INFICON General Manager – Service Tools Two Technology Place East Syracuse, NY USA 13057 EU Authorized Representative INFICON GmbH 50968 Köln, Bonner Str. 498

ANY QUESTIONS RELATIVE TO THIS DECLARATION OR TO THE SAFETY OF INFICON'S PRODUCTS SHOULD BE DIRECTED, IN WRITING, TO THE AUTHORIZED REPRESENTATIVE AT THE ABOVE ADDRESS.